

CE

Declaration of conformity



**Acoustic Precise
Measurement**

APM Automation Solutions Ltd. 24 Habarzel St., Tel-Aviv 69710, Israel
Tel: 972-3-6488891 / Fax: 972-3-6488892
e-mail: info@apm-solutions.com / website: www.apm-solutions.com

our sole responsibility that our product:

3DLevelScanner S, M, MV
4...20mA/HART/RS485

to which this declaration relates is in conformity
with the following standards

EMC: EN61326:1997 + A1: 1998 + A2: 2001 + A3: 2003
Safety: IEC / EN 61010 – 1:2001

following the provision
of Directives

73/23/EEC
89/336/EEC

Tel- Aviv 30/10/08

Ofir Perl



HERMON LABORATORIES



Certificate of Conformity
with EMC Directive 2004/108/EC, Article 5

N° APMEMC_EN.17614C_rev3

Product definition : 3DLevelScanner
Product types : M, S, MV 4...20mA/HART/RS485
Applicant / Manufacturer : APM Automation Solutions Ltd.
Address : 24 Habarzel Street, Tel Aviv 69710, Israel
Telephone number : +972 3648 8891
Fax number : +972 3648 8892

This Certificate of Conformity has been granted to the applicant based on the results of tests and measurements, performed by Hermon Laboratories on a sample of the above-mentioned product from December 19, 2006 to January 12, 2007.

The tested product satisfies the requirements of:

EN 61326:1997 + A1:1998 + A1:2001 + A3:2003 standard, harmonized under EMC Directive 2004/108/EC Article 6(2), for equipment intended for use in industrial locations, complies with class A limits for radiated emissions.

Measurements/tests results are contained in the test report:
APMEMC_EN.17614_rev3.

Michael Nikishin, EMC and Radio Group Leader
Hermon Laboratories Ltd.

October 29, 2008

The comments in the associated (if applicable) test report/s shall be taken into account and used in conjunction with this certificate. The "A2LA Accredited" symbol endorsement applies only to the tests and calibrations that are listed in the scope of Hermon Laboratories accreditation.

P.O. Box 23 Binyamina 30500, Israel
Tel. +972-4-6288001
Fax +972-4-6288277
www.hermonlabs.com



Certificate of Conformity
with
EN/IEC 61010-1 standard

N° APMSAF_17614_rev1C

Product definition: 3DLevelScanner
Model: M, S, MV 4...20mA/HART/RS485
Ratings and principal characteristics.....: 20-36Vdc, Max 20mA
Applicant/Manufacturer: APM Automation Solutions Ltd
Address.....: 24 Habarzel Street, Tel Aviv 69710, Israel
Telephone number.....: +972 3648 8891
Fax number.....: +972 3648 8892

This Certificate of Conformity has been granted to the applicant based on the results of tests and measurements, performed by Hermon Laboratories on a sample of the above-mentioned product from 20 to 30 December, 2006.

The tested product satisfies the requirements of:
EN 61010-1:2001

Measurements/tests results are contained in the test report:
APMSAF_EN.17614_rev1

Michael Brun
Product Safety Group Manager
Hermon Laboratories Ltd.

26 November 2008

The comments in the associated (if applicable) test report/s shall be taken into account and used in conjunction with this certificate. The "A2LA Accredited" symbol endorsement applies only to the tests and calibrations that are listed in the scope of Hermon Laboratories accreditation.