



EMC Test Certificate

| | | |
|-----------------------|-------------|----------------------|
| Certificate No | Page | Date of Issue |
| K104550.00 | 1 | 04 April 2011 |

Applicant **APM Automation Solutions Ltd.**
Atidim High-Tech Industrial Park, Building 2, P.O.B 58171,
Tel Aviv 61580, Israel



Tested to

| | |
|-------------------------|--|
| EN 61326-1: 2006 | |
| CISPR 11: 2003, Class A | |
| IEC 61000-4-2: 2001 | Air Discharge, 8kV |
| | Contact Discharge, 4kV |
| IEC 61000-4-3: 2002 | (80-1000 MHz), 1V/m; (1.4-2 GHz), 1V/m |
| | (2.0-2.7 GHz) 1V/m |
| | 80% A.M. by 1kHz |
| IEC 61000-4-4: 2004 | Power Lines: 1kV Signal Lines 0.5kV |
| IEC 61000-4-6: 2003 | (0.15-80 MHz) 1VRMS, 80% A.M. by 1kHz |
| | Power and Signal Lines |

Note: The above list of tests was performed according to the customer's request.

Certified Product

| | |
|-------------|--------------------|
| E.U.T. | 3D LevelScanner II |
| Model: | ASM0000092-0 "S" |
| Serial No.: | 25100172 |

This is to certify that the product specified herein has been tested and the test results were found to be compliant with the requirements noted above.

Signature:
 S. Sasson
 EMC Test Engineer

Signature:
 I. Raz
 EMC Laboratory Manager

ITL091 Rev 1.7 26/06/07

